
Midmark M11 - v1.0.0
Total Cycles: 3709

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2023 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 163.1 F 2.3
4:00 181.7 F 6.0
6:00 201.0 F 10.4
8:00 222.8 F 5.3
10:00 228.4 F 9.3
12:00 238.1 F 10.3
14:00 241.0 F 11.2
16:00 237.3 F 9.0
18:00 250.7 F 15.4
20:00 261.8 F 21.7
22:00 271.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.1 F 28.5
2:00 272.7 F 28.8
3:00 272.5 F 29.1
4:00 271.6 F 28.5

Min 271.6 F 28.3
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 223.1 F 0.0
10:00 213.9 F 0.0
15:00 208.7 F 0.0
20:00 195.9 F 0.0
25:00 195.6 F 0.0
30:00 202.5 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:02:32
CYCLE COMPLETE

08/31/2023 11:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____