

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4928

Name: RITTER 1  
Sterilizer ID: V2178198  
09/20/2024 02:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.0 F 0.0  
2:00 161.4 F 0.3  
4:00 189.2 F 2.3  
6:00 209.4 F 6.9  
8:00 226.2 F 12.3  
10:00 229.7 F 7.4  
12:00 241.0 F 12.6  
14:00 239.7 F 10.3  
16:00 235.8 F 8.3  
18:00 251.2 F 15.7  
20:00 263.2 F 22.5  
22:00 272.3 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.8  
2:00 271.5 F 28.1  
3:00 272.5 F 28.9  
4:00 271.8 F 28.1

Min 271.4 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.7 F 0.0  
5:00 238.7 F 0.0  
10:00 201.0 F 0.0  
15:00 191.8 F 0.0  
20:00 229.1 F 0.0  
25:00 225.4 F 0.0  
30:00 211.8 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 22:12  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:00:46  
CYCLE COMPLETE

09/20/2024 03:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_