
Midmark M11 - v1.0.5
Total Cycles: 1728

Name: RITTER2
Sterilizer ID: V2313228
11/15/2022 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.8 F 0.0
2:00 175.8 F 0.6
4:00 191.0 F 4.1
6:00 207.5 F 9.2
8:00 223.4 F 8.3
10:00 234.4 F 10.9
12:00 236.2 F 8.7
14:00 240.7 F 10.0
16:00 248.3 F 14.8
18:00 261.0 F 22.0
20:00 272.3 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.8 F 29.7
2:00 272.0 F 29.0
3:00 272.2 F 29.6
4:00 272.7 F 29.6

Min 271.6 F 28.8
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 200.9 F 0.0
10:00 214.7 F 0.0
15:00 223.2 F 0.0
20:00 225.1 F 0.0
25:00 225.8 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

11/15/2022 03:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____