
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004270
13:09 30/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:31
STERILIZING 4:46
136.4 C 222kPa 4:46
136.3 C 224kPa 5:17
136.2 C 223kPa 5:47
136.3 C 224kPa 6:17
136.4 C 224kPa 6:47
136.3 C 222kPa 7:17
136.3 C 223kPa 7:47
136.4 C 225kPa 8:17
136.4 C 225kPa 8:17
VENTING 8:17
AIR DRYING 9:07
CYCLE COMPLETE 11:27

Digital Signature #
17062AD7F3349FC2
