
Midmark M11 - v1.0.0
Total Cycles: 1725

Name: RITTER 1
Sterilizer ID: V2178198
03/16/2022 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.5 F 0.0
2:00 179.6 F 3.0
4:00 185.1 F 7.3
6:00 210.8 F 12.9
8:00 231.5 F 7.9
10:00 239.1 F 13.4
12:00 241.8 F 11.0
14:00 238.4 F 9.4
16:00 252.6 F 16.4
18:00 263.4 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.3 F 28.8
2:00 272.5 F 29.2
3:00 271.6 F 28.4
4:00 272.7 F 29.2

Min 271.3 F 28.4
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.2 F 0.0
5:00 219.4 F 0.0
10:00 225.9 F 0.0
15:00 226.4 F 0.0
20:00 231.0 F 0.0
25:00 229.6 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

03/16/2022 03:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____