

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000509  
9:49 21/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:45  
PRESSURIZING 3:48  
STERILIZING 5:01  
136.0 C 227kPa 5:01  
135.7 C 227kPa 5:31  
135.9 C 228kPa 6:01  
136.2 C 230kPa 6:31  
136.2 C 231kPa 7:01  
136.3 C 231kPa 7:31  
136.4 C 232kPa 8:01  
136.4 C 232kPa 8:31  
136.4 C 232kPa 8:31  
VENTING 8:31  
AIR DRYING 9:26  
CYCLE COMPLETE 24:27  
-----

Digital Signature #  
B16C04FCCA09007E  
-----