
Midmark M11 - v1.0.5
Total Cycles: 2440

Name: RITTER2
Sterilizer ID: V2313228
05/31/2023 07:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 76.4 F 0.0
4:00 80.8 F 0.0
6:00 109.7 F 0.0
8:00 143.7 F 1.5
10:00 172.7 F 4.9
12:00 195.5 F 8.5
14:00 208.1 F 12.1
16:00 215.3 F 5.7
18:00 227.0 F 9.7
20:00 235.1 F 10.6
22:00 237.8 F 10.8
24:00 235.5 F 8.6
26:00 248.7 F 15.1
28:00 259.8 F 21.3
30:00 270.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.7 F 29.7
2:00 271.7 F 28.9
3:00 271.9 F 29.6
4:00 272.5 F 29.5

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 194.2 F 0.0
10:00 205.7 F 0.0
15:00 210.2 F 0.0
20:00 208.8 F 0.0
25:00 212.5 F 0.0
30:00 213.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:56
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

05/31/2023 08:21 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____