
Midmark M11 - v1.0.5
Total Cycles: 4117

Name: RITTER2
Sterilizer ID: V2313228
11/14/2024 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.5 F 0.0
4:00 106.9 F 0.0
6:00 98.1 F 0.1
8:00 144.0 F 0.2
10:00 176.2 F 0.5
12:00 196.0 F 0.6
14:00 212.3 F 3.4
16:00 220.5 F 6.2
18:00 228.9 F 9.5
20:00 237.9 F 13.2
22:00 236.1 F 9.7
24:00 240.4 F 10.4
26:00 245.2 F 13.1
28:00 243.6 F 12.2
30:00 255.3 F 18.3
32:00 265.5 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.1
2:00 271.6 F 28.9
3:00 272.4 F 28.9
4:00 271.6 F 28.9

Min 271.4 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 207.3 F 0.0
10:00 190.6 F 0.0
15:00 183.1 F 0.0
20:00 182.0 F 0.0
25:00 184.0 F 0.0
30:00 193.7 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 33:54
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:12:23
CYCLE COMPLETE

11/14/2024 11:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____