
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000891
10:58 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:04
STERILIZING 3:09
136.2 C 220kPa 3:09
136.3 C 232kPa 3:39
136.3 C 231kPa 4:09
136.3 C 231kPa 4:39
136.2 C 230kPa 5:09
136.4 C 232kPa 5:39
136.3 C 231kPa 6:09
136.4 C 232kPa 6:39
136.4 C 232kPa 6:39
VENTING 6:39
AIR DRYING 7:32
CYCLE COMPLETE 22:32

Digital Signature #
52013B299CC52226
