

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2492

Name: RITTER 1  
Sterilizer ID: V2178198  
10/26/2022 11:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.7 F 0.0  
2:00 186.3 F 4.3  
4:00 192.4 F 8.8  
6:00 217.1 F 7.4  
8:00 225.0 F 8.6  
10:00 236.2 F 11.2  
12:00 239.5 F 11.2  
14:00 237.9 F 9.6  
16:00 251.5 F 16.4  
18:00 263.4 F 23.4  
20:00 271.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.6 F 29.5  
2:00 272.0 F 28.9  
3:00 271.6 F 29.0  
4:00 272.6 F 29.5  
  
Min 271.3 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 217.3 F 0.0  
10:00 230.8 F 0.0  
15:00 229.6 F 0.0  
20:00 226.6 F 0.0  
25:00 224.2 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 20:10  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:24  
CYCLE COMPLETE

10/26/2022 12:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_