
Midmark M11 - v1.0.0
Total Cycles: 3651

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2023 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 80.4 F 0.0
4:00 85.8 F 0.0
6:00 117.2 F 1.8
8:00 148.2 F 4.3
10:00 172.1 F 7.3
12:00 184.9 F 10.5
14:00 201.7 F 6.6
16:00 214.2 F 5.1
18:00 224.2 F 8.6
20:00 234.7 F 12.6
22:00 234.6 F 8.7
24:00 242.5 F 11.6
26:00 244.2 F 12.3
28:00 256.2 F 18.5
30:00 266.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.1 F 28.3
2:00 272.3 F 28.4
3:00 272.5 F 28.6
4:00 272.7 F 28.9

Min 271.6 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 161.5 F 0.0
5:00 198.6 F 0.0
10:00 195.2 F 0.0
15:00 202.2 F 0.0
20:00 198.7 F 0.0
25:00 198.3 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:22
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:25
CYCLE COMPLETE

08/15/2023 09:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____