
Midmark M11 - v1.0.5
Total Cycles: 1862

Name: RITTER2
Sterilizer ID: V2313228
12/20/2022 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.5 F 0.0
2:00 189.3 F 3.8
4:00 201.4 F 8.8
6:00 214.1 F 7.2
8:00 229.4 F 9.4
10:00 233.3 F 9.1
12:00 243.3 F 13.1
14:00 243.3 F 12.3
16:00 256.6 F 19.3
18:00 268.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.3 F 29.5
2:00 271.5 F 28.8
3:00 272.0 F 29.7
4:00 272.2 F 29.4

Min 271.1 F 28.7
Max 272.4 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 224.9 F 0.0
10:00 241.6 F 0.0
15:00 237.6 F 0.0
20:00 227.8 F 0.0
25:00 230.4 F 0.0
30:00 226.8 F 0.0

Drying Complete

FILLING: 1:29
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

12/20/2022 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____