
Midmark M11 - v1.0.5
Total Cycles: 3065

Name: RITTER2
Sterilizer ID: V2313228
10/30/2023 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.1 F 0.0
2:00 122.9 F 0.6
4:00 149.0 F 3.1
6:00 173.1 F 6.6
8:00 192.7 F 10.7
10:00 214.3 F 5.3
12:00 220.6 F 9.2
14:00 232.4 F 10.9
16:00 236.4 F 11.2
18:00 237.5 F 9.6
20:00 250.8 F 16.0
22:00 262.2 F 22.6
24:00 272.9 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.9 F 29.9
2:00 271.0 F 29.5
3:00 273.0 F 30.2
4:00 272.5 F 30.5

Min 271.0 F 29.3
Max 273.5 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 237.6 F 0.0
10:00 225.1 F 0.0
15:00 182.0 F 0.0
20:00 178.7 F 0.0
25:00 175.6 F 0.0
30:00 173.7 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:09
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:03:44
CYCLE COMPLETE

10/30/2023 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____