
Midmark M11 - v1.0.5
Total Cycles: 1722

Name: RITTER2
Sterilizer ID: V2313228
11/15/2022 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.2 F 0.0
2:00 80.5 F 0.0
4:00 107.5 F 0.0
6:00 142.7 F 0.0
8:00 161.4 F 0.1
10:00 208.5 F 0.3
12:00 216.4 F 3.3
14:00 224.9 F 6.7
16:00 235.9 F 10.4
18:00 235.2 F 8.9
20:00 242.9 F 11.8
22:00 239.1 F 9.8
24:00 237.6 F 9.0
26:00 250.1 F 15.1
28:00 261.6 F 21.7
30:00 272.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.5 F 28.8
2:00 271.2 F 28.4
3:00 272.3 F 28.8
4:00 271.2 F 28.2

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 240.5 F 0.0
10:00 221.6 F 0.0
15:00 220.0 F 0.0
20:00 210.7 F 0.0
25:00 210.9 F 0.0
Dry Cycle Aborted

FILLING: 1:00
HEATING: 30:29
STERILIZING: 4:00
VENTING: 3:07
DRYING: 28:21
TOTAL CYCLE: 01:06:57

11/15/2022 08:58 AM
ACCEPTED: X
REJECTED:
Operator: U02