
Midmark M11 - v1.0.5
Total Cycles: 4193

Name: RITTER2
Sterilizer ID: V2313228
12/06/2024 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.0 F 0.0
2:00 148.7 F 0.2
4:00 165.2 F 0.9
6:00 205.1 F 4.8
8:00 208.5 F 9.1
10:00 227.5 F 11.6
12:00 231.6 F 9.1
14:00 237.2 F 11.1
16:00 241.5 F 11.9
18:00 240.0 F 10.6
20:00 252.7 F 16.9
22:00 263.4 F 23.2
24:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.2 F 29.3
2:00 273.2 F 29.7
3:00 272.1 F 29.9
4:00 271.5 F 29.2

Min 271.3 F 29.2
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 194.6 F 0.0
10:00 189.1 F 0.0
15:00 189.0 F 0.0
20:00 187.5 F 0.0
25:00 186.5 F 0.0
30:00 185.4 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 24:57
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:54
CYCLE COMPLETE

12/06/2024 01:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____