
Midmark M11 - v1.0.0
Total Cycles: 1960

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2022 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.7 F 0.0
2:00 176.3 F 4.0
4:00 189.1 F 8.3
6:00 215.6 F 12.6
8:00 224.6 F 8.2
10:00 235.3 F 13.1
12:00 236.3 F 10.3
14:00 235.8 F 8.0
16:00 248.7 F 14.9
18:00 260.5 F 21.6
20:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.6 F 30.1
2:00 272.6 F 29.6
3:00 271.8 F 29.6
4:00 272.7 F 30.0

Min 271.0 F 28.7
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 202.8 F 0.0
10:00 214.2 F 0.0
15:00 211.4 F 0.0
20:00 213.2 F 0.0
25:00 216.8 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

05/26/2022 03:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____