
Midmark M11 - v1.0.0
Total Cycles: 1086

Name: RITTER 1
Sterilizer ID: V2178198
08/10/2021 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.9 F 0.0
2:00 142.3 F 1.1
4:00 157.4 F 3.6
6:00 181.3 F 7.4
8:00 200.7 F 11.9
10:00 217.1 F 5.3
12:00 230.5 F 9.9
14:00 233.5 F 8.8
16:00 241.6 F 11.9
18:00 239.2 F 10.1
20:00 252.0 F 16.3
22:00 263.0 F 22.7
24:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.8
2:00 271.6 F 28.5
3:00 272.5 F 28.8
4:00 271.4 F 28.3

Min 271.4 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 206.5 F 0.0
10:00 212.2 F 0.0
15:00 217.8 F 0.0
20:00 213.1 F 0.0
25:00 209.0 F 0.0
30:00 206.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:02:42
CYCLE COMPLETE

08/10/2021 12:00 PM
ACCEPTED: X
REJECTED:
Operator: U01