
Midmark M11 - v1.0.5
Total Cycles: 1112

Name: RITTER2
Sterilizer ID: V2313228
05/25/2022 07:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 75.9 F 0.0
4:00 100.0 F 0.0
6:00 105.6 F 0.2
8:00 141.6 F 3.1
10:00 168.0 F 6.4
12:00 186.5 F 10.0
14:00 207.6 F 8.9
16:00 218.1 F 6.7
18:00 229.9 F 10.7
20:00 231.6 F 7.8
22:00 241.2 F 12.2
24:00 239.5 F 10.5
26:00 252.0 F 16.7
28:00 262.7 F 22.9
30:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.1 F 29.0
2:00 272.5 F 29.5
3:00 271.5 F 29.2
4:00 272.0 F 29.0

Min 271.4 F 28.8
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 201.7 F 0.0
10:00 200.6 F 0.0
15:00 200.0 F 0.0
20:00 201.3 F 0.0
25:00 203.7 F 0.0
30:00 206.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:19
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:08:10
CYCLE COMPLETE

05/25/2022 08:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____