

-----  
Midmark M11 - v1.0.5  
Total Cycles: 947

Name: RITTER2  
Sterilizer ID: V2313228  
04/01/2022 02:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.4 F 0.0  
2:00 196.0 F 1.7  
4:00 204.2 F 7.5  
6:00 224.1 F 13.5  
8:00 232.8 F 10.5  
10:00 234.5 F 9.1  
12:00 237.6 F 9.4  
14:00 250.7 F 16.3  
16:00 263.9 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.8 F 29.7  
2:00 271.6 F 29.1  
3:00 272.8 F 29.9  
4:00 272.1 F 29.2  
  
Min 271.5 F 28.9  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
5:00 222.2 F 0.0  
10:00 237.3 F 0.0  
15:00 232.8 F 0.0  
20:00 223.8 F 0.0  
25:00 221.7 F 0.0  
30:00 220.9 F 0.0  
Drying Complete

FILLING: 2:54  
HEATING: 17:50  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 00:57:46  
CYCLE COMPLETE

04/04/2022 07:00 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02