
Midmark M11 - v1.0.0
Total Cycles: 4268

Name: RITTER 1
Sterilizer ID: V2178198
02/05/2024 10:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.1 F 0.0
2:00 170.6 F 1.8
4:00 169.5 F 5.0
6:00 200.3 F 8.5
8:00 206.0 F 13.3
10:00 219.9 F 6.6
12:00 233.4 F 11.2
14:00 233.1 F 7.7
16:00 243.0 F 13.4
18:00 242.3 F 11.7
20:00 254.2 F 17.6
22:00 264.4 F 23.8
24:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.5 F 28.7
2:00 272.8 F 29.2
3:00 271.9 F 28.8
4:00 271.2 F 28.3

Min 271.1 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 223.3 F 0.0
10:00 224.6 F 0.0
15:00 213.0 F 0.0
20:00 211.4 F 0.0
25:00 212.4 F 0.0
30:00 215.3 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

02/05/2024 11:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____