

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2

Name: RITTER2  
Sterilizer ID: V2313228  
04/21/2021 07:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.5 F 0.0  
2:00 201.4 F 5.8  
4:00 224.0 F 10.7  
6:00 241.3 F 13.4  
8:00 244.1 F 12.3  
10:00 250.7 F 15.9  
12:00 265.6 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 271.9 F 28.9  
2:00 272.0 F 29.0  
3:00 272.1 F 29.0  
4:00 272.0 F 28.9

Min 271.8 F 28.6  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
Dry Cycle Aborted

FILLING: 3:01  
HEATING: 13:22  
STERILIZING: 4:00  
VENTING: 2:49  
DRYING: 4:17  
TOTAL CYCLE: 00:27:29

04/21/2021 07:58 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01