

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2160

Name: RITTER2  
Sterilizer ID: V2313228  
03/20/2023 09:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.2 F 0.0  
2:00 148.5 F 1.8  
4:00 160.4 F 4.6  
6:00 184.5 F 9.1  
8:00 215.0 F 8.3  
10:00 221.4 F 8.9  
12:00 232.9 F 10.4  
14:00 238.2 F 11.5  
16:00 236.9 F 9.7  
18:00 250.2 F 16.1  
20:00 261.8 F 22.8  
22:00 272.4 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.4  
1:00 272.8 F 30.1  
2:00 271.5 F 29.3  
3:00 271.9 F 30.0  
4:00 272.4 F 29.8

Min 271.1 F 29.2  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
5:00 226.1 F 0.0  
10:00 237.9 F 0.0  
15:00 223.2 F 0.0  
20:00 220.1 F 0.0  
25:00 219.8 F 0.0  
30:00 217.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:28  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:43  
CYCLE COMPLETE

03/20/2023 10:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_