
Midmark M11 - v1.0.0
Total Cycles: 5382

Name: RITTER 1
Sterilizer ID: V2178198
03/03/2025 12:10 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.4 F 0.0
2:00 140.2 F 1.5
4:00 145.5 F 4.4
6:00 173.6 F 8.7
8:00 194.1 F 7.8
10:00 217.1 F 6.4
12:00 232.2 F 11.6
14:00 232.2 F 7.9
16:00 242.6 F 11.3
18:00 245.1 F 12.7
20:00 258.2 F 19.6
22:00 269.8 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2

C534: STERILIZE MODE
STEAM TEMP LOW
CHAMBER TEMP 270F
CHAMBER PRESS 21.7PSI
ELAPSED TIME 00:32
** ITEMS NOT STERILE **

03/03/2025 12:34 PM