
Midmark M11 - v1.0.5
Total Cycles: 3757

Name: RITTER2
Sterilizer ID: V2313228
07/05/2024 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.9 F 0.0
2:00 120.9 F 1.2
4:00 136.5 F 3.7
6:00 170.2 F 7.1
8:00 193.4 F 11.4
10:00 212.7 F 4.9
12:00 222.9 F 9.3
14:00 232.5 F 11.0
16:00 236.5 F 10.9
18:00 235.0 F 9.0
20:00 248.9 F 15.3
22:00 259.9 F 21.6
24:00 270.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.2 F 29.7
2:00 271.3 F 29.7
3:00 272.0 F 29.6
4:00 272.0 F 30.3

Min 270.9 F 29.2
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 209.4 F 0.0
10:00 193.1 F 0.0
15:00 189.4 F 0.0
20:00 186.9 F 0.0
25:00 186.1 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:52
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:27
CYCLE COMPLETE

07/05/2024 02:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____