

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007745  
10:29 14/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:35  
STERILIZING 3:36  
136.0 C 220kPa 3:36  
136.2 C 221kPa 4:06  
136.4 C 224kPa 4:36  
136.2 C 221kPa 5:06  
136.2 C 223kPa 5:36  
136.4 C 220kPa 6:06  
136.4 C 221kPa 6:36  
136.1 C 219kPa 7:06  
136.1 C 219kPa 7:06  
VENTING 7:06  
AIR DRYING 7:58  
CYCLE COMPLETE 37:58

-----  
Digital Signature #  
27EE4480F9A35814  
-----