

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001054  
10:58 04/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:54  
STERILIZING 4:28  
136.4 C 222kPa 4:28  
136.3 C 224kPa 4:58  
136.4 C 222kPa 5:28  
136.0 C 221kPa 5:58  
136.2 C 222kPa 6:28  
136.3 C 223kPa 6:58  
136.4 C 224kPa 7:28  
136.3 C 223kPa 7:58  
136.3 C 223kPa 7:58  
VENTING 7:58  
AIR DRYING 8:49  
CYCLE COMPLETE 23:49  
-----

Digital Signature #  
9FFFBA09C721738E  
-----