
Midmark M11 - v1.0.0
Total Cycles: 4491

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2024 09:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.7 F 0.0
2:00 172.6 F 2.1
4:00 172.9 F 5.3
6:00 189.6 F 9.4
8:00 207.1 F 9.3
10:00 224.8 F 7.2
12:00 234.1 F 11.8
14:00 233.0 F 8.3
16:00 242.9 F 12.1
18:00 244.1 F 12.2
20:00 255.9 F 18.3
22:00 266.4 F 24.8
24:00 272.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.5 F 28.9
2:00 272.2 F 29.3
3:00 271.6 F 28.6
4:00 272.5 F 29.1

Min 271.4 F 28.6
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 217.4 F 0.0
10:00 220.5 F 0.0
15:00 230.9 F 0.0
20:00 234.5 F 0.0
25:00 225.7 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:24
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:10
CYCLE COMPLETE

04/25/2024 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____