
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000912
12:42 07/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:01
PRESSURIZING 4:03
STERILIZING 5:10
136.0 C 228kPa 5:10
136.3 C 232kPa 5:40
136.3 C 231kPa 6:10
136.4 C 232kPa 6:40
136.1 C 231kPa 7:10
136.3 C 232kPa 7:40
136.1 C 230kPa 8:10
136.3 C 232kPa 8:40
136.3 C 232kPa 8:40
VENTING 8:40
AIR DRYING 9:32
CYCLE COMPLETE 24:32

Digital Signature #
30802AFC7842B675
