

-----  
Midmark M11 - v1.0.5  
Total Cycles: 651

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2021 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.3 F 0.0  
2:00 166.0 F 2.6  
4:00 188.6 F 7.0  
6:00 209.9 F 11.8  
8:00 224.6 F 6.7  
10:00 236.6 F 11.9  
12:00 236.1 F 9.6  
14:00 238.2 F 9.5  
16:00 249.4 F 15.3  
18:00 261.5 F 22.1  
20:00 272.7 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.5 F 29.3  
2:00 271.2 F 28.8  
3:00 272.6 F 29.4  
4:00 271.5 F 28.7

Min 271.2 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.1 F 0.0  
5:00 231.8 F 0.0  
10:00 229.2 F 0.0  
15:00 191.7 F 0.0  
20:00 185.7 F 0.0  
25:00 182.9 F 0.0  
30:00 181.8 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:17  
CYCLE COMPLETE

12/20/2021 11:12 AM  
ACCEPTED: X  
REJECTED:  
Operator: U04