
Midmark M11 - v1.0.5
Total Cycles: 837

Name: RITTER2
Sterilizer ID: V2313228
02/24/2022 12:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.6 F 0.0
2:00 193.8 F 3.5
4:00 199.2 F 9.1
6:00 218.0 F 7.3
8:00 234.7 F 10.2
10:00 232.5 F 8.0
12:00 242.3 F 11.5
14:00 246.5 F 13.6
16:00 259.3 F 20.7
18:00 270.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.7 F 29.6
2:00 271.9 F 28.8
3:00 273.0 F 29.8
4:00 272.6 F 29.2

Min 271.7 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 236.7 F 0.0
10:00 229.1 F 0.0
15:00 226.5 F 0.0
20:00 228.4 F 0.0
25:00 224.0 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 2:45
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

02/24/2022 01:48 PM
ACCEPTED: X
REJECTED:
Operator: U02