
Midmark M11 - v1.0.0
Total Cycles: 2747

Name: RITTER 1
Sterilizer ID: V2178198
01/06/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.2 F 0.0
2:00 192.4 F 0.0
4:00 200.9 F 4.0
6:00 215.5 F 9.2
8:00 229.1 F 9.4
10:00 237.0 F 10.4
12:00 234.3 F 8.0
14:00 242.9 F 11.5
16:00 246.4 F 13.3
18:00 258.9 F 20.0
20:00 269.9 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.7
2:00 271.5 F 28.1
3:00 272.6 F 28.9
4:00 271.7 F 28.1

Min 271.4 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 233.5 F 0.0
10:00 225.5 F 0.0
15:00 228.3 F 0.0
20:00 225.8 F 0.0
25:00 226.6 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

01/06/2023 11:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____