

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1514

Name: RITTER2  
Sterilizer ID: V2313228  
09/21/2022 12:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.4 F 0.0  
2:00 121.9 F 0.0  
4:00 136.2 F 1.8  
6:00 168.3 F 5.7  
8:00 190.5 F 10.2  
10:00 214.2 F 6.3  
12:00 225.2 F 8.8  
14:00 237.7 F 12.7  
16:00 238.1 F 10.6  
18:00 235.6 F 8.4  
20:00 247.7 F 14.6  
22:00 258.5 F 20.7  
24:00 269.1 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 272.3 F 29.6  
2:00 271.3 F 29.4  
3:00 272.3 F 29.6  
4:00 271.1 F 29.2

Min 271.1 F 29.0  
Max 272.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 228.2 F 0.0  
10:00 224.3 F 0.0  
15:00 225.0 F 0.0  
20:00 228.6 F 0.0  
25:00 226.7 F 0.0  
30:00 221.5 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 25:04  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:03:10  
CYCLE COMPLETE

09/21/2022 01:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_