
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000835
14:28 30/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:31
STERILIZING 4:56
136.0 C 220kPa 4:56
136.3 C 230kPa 5:27
136.2 C 231kPa 5:57
136.3 C 232kPa 6:27
136.4 C 232kPa 6:57
136.4 C 231kPa 7:27
136.5 C 231kPa 7:57
136.4 C 232kPa 8:27
136.4 C 232kPa 8:27
VENTING 8:27
AIR DRYING 9:19
CYCLE COMPLETE 24:19

Digital Signature #
81F89143DE756EB0
