
Midmark M11 - v1.0.5
Total Cycles: 3132

Name: RITTER2
Sterilizer ID: V2313228
11/16/2023 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.6 F 0.0
2:00 168.3 F 1.4
4:00 175.5 F 5.3
6:00 195.4 F 10.0
8:00 217.1 F 6.7
10:00 229.4 F 10.1
12:00 231.4 F 8.0
14:00 240.8 F 13.4
16:00 243.6 F 12.6
18:00 256.2 F 19.3
20:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.8 F 29.9
2:00 271.5 F 29.2
3:00 271.7 F 29.8
4:00 272.5 F 29.7

Min 271.0 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 209.4 F 0.0
10:00 217.1 F 0.0
15:00 206.5 F 0.0
20:00 199.6 F 0.0
25:00 198.1 F 0.0
30:00 197.9 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

11/16/2023 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____