
Midmark M11 - v1.0.0
Total Cycles: 2796

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2023 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.5 F 0.0
2:00 185.7 F 2.9
4:00 193.0 F 6.9
6:00 208.5 F 12.0
8:00 220.1 F 6.0
10:00 231.9 F 11.2
12:00 233.7 F 8.4
14:00 239.6 F 10.4
16:00 246.7 F 13.5
18:00 259.1 F 20.3
20:00 270.3 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.3 F 28.9
2:00 271.7 F 28.8
3:00 272.1 F 28.8
4:00 271.5 F 28.7

Min 271.4 F 28.4
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 230.4 F 0.0
10:00 227.7 F 0.0
15:00 228.4 F 0.0
20:00 228.6 F 0.0
25:00 228.7 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

01/19/2023 11:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____