

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3737

Name: RITTER2  
Sterilizer ID: V2313228  
06/26/2024 01:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.7 F 0.0  
2:00 137.7 F 1.9  
4:00 149.6 F 4.8  
6:00 174.8 F 8.5  
8:00 193.7 F 13.0  
10:00 210.3 F 6.8  
12:00 222.6 F 11.4  
14:00 228.8 F 8.0  
16:00 241.7 F 13.5  
18:00 241.8 F 12.3  
20:00 254.3 F 18.5  
22:00 264.9 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.5  
1:00 272.9 F 30.3  
2:00 272.8 F 30.3  
3:00 272.7 F 30.4  
4:00 272.2 F 30.4

Min 270.9 F 29.4  
Max 273.0 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
5:00 205.9 F 0.0  
10:00 200.5 F 0.0  
15:00 189.8 F 0.0  
20:00 186.8 F 0.0  
25:00 184.7 F 0.0  
30:00 183.3 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 23:45  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:02:35  
CYCLE COMPLETE

06/26/2024 02:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_