

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3142

Name: RITTER2  
Sterilizer ID: V2313228  
11/20/2023 09:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.3 F 0.0  
2:00 81.6 F 0.0  
4:00 76.7 F 0.3  
6:00 102.3 F 1.5  
8:00 141.5 F 4.0  
10:00 169.7 F 6.9  
12:00 203.2 F 10.5  
14:00 213.9 F 6.1  
16:00 212.4 F 8.0  
18:00 224.2 F 12.0  
20:00 231.0 F 7.5  
22:00 239.6 F 12.7  
24:00 239.5 F 10.5  
26:00 251.6 F 16.5  
28:00 262.4 F 22.8  
30:00 272.9 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 271.6 F 29.4  
2:00 271.7 F 29.5  
3:00 271.6 F 29.4  
4:00 271.5 F 29.3

Min 271.4 F 29.2  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.6 F 0.0  
5:00 195.7 F 0.0  
10:00 184.6 F 0.0  
15:00 184.1 F 0.0  
20:00 182.6 F 0.0  
25:00 181.9 F 0.0  
30:00 181.1 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:03  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:09:10  
CYCLE COMPLETE

11/20/2023 11:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_