
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000905
12:14 04/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:52
STERILIZING 4:27
136.4 C 233kPa 4:27
136.3 C 231kPa 4:57
136.4 C 233kPa 5:27
136.4 C 231kPa 5:57
136.2 C 231kPa 6:27
136.4 C 233kPa 6:57
136.4 C 231kPa 7:27
136.2 C 231kPa 7:57
136.2 C 231kPa 7:57
VENTING 7:57
AIR DRYING 8:53
CYCLE COMPLETE 12:35

Digital Signature #
D8F612159453C72C
