
Midmark M11 - v1.0.5
Total Cycles: 4210

Name: RITTER2
Sterilizer ID: V2313228
12/13/2024 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.4 F 0.0
4:00 77.4 F 0.0
6:00 99.1 F 0.0
8:00 142.4 F 0.1
10:00 173.1 F 2.1
12:00 194.7 F 5.3
14:00 204.8 F 8.6
16:00 217.2 F 12.3
18:00 225.3 F 6.7
20:00 235.1 F 11.1
22:00 233.5 F 7.6
24:00 245.5 F 13.5
26:00 243.5 F 12.0
28:00 255.2 F 18.1
30:00 265.8 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.1
2:00 271.3 F 28.5
3:00 272.5 F 29.2
4:00 271.6 F 28.4

Min 271.3 F 28.3
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 217.3 F 0.0
10:00 186.6 F 0.0
15:00 185.0 F 0.0
20:00 183.3 F 0.0
25:00 183.0 F 0.0
30:00 183.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:46
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:53
CYCLE COMPLETE

12/13/2024 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____