
Midmark M11 - v1.0.0
Total Cycles: 4927

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2024 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.0 F 0.0
2:00 150.8 F 0.1
4:00 173.0 F 0.5
6:00 192.5 F 3.2
8:00 213.0 F 6.6
10:00 221.3 F 10.9
12:00 225.5 F 6.5
14:00 235.4 F 10.4
16:00 234.7 F 8.3
18:00 245.1 F 13.2
20:00 243.1 F 11.8
22:00 254.8 F 17.7
24:00 265.0 F 23.8
26:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.1 F 28.7
2:00 272.2 F 28.6
3:00 271.6 F 28.2
4:00 271.4 F 28.3

Min 271.3 F 28.0
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 209.2 F 0.0
10:00 219.9 F 0.0
15:00 225.1 F 0.0
20:00 237.4 F 0.0
25:00 230.0 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:08
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:04:58
CYCLE COMPLETE

09/20/2024 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____