

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 001426  
12:20 09/13/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:03  
PRESSURIZING 6:00  
STERILIZING 7:46  
136.0 C 220kPa 7:46  
136.5 C 223kPa 8:16  
136.4 C 222kPa 8:46  
136.4 C 223kPa 9:16  
136.1 C 221kPa 9:46  
136.4 C 219kPa 10:16  
136.4 C 217kPa 10:46  
136.4 C 222kPa 11:16  
VENTING 11:16  
AIR DRYING 12:15  
CYCLE COMPLETE 36:52

-----  
Digital Signature #  
B60BD4B285608EC4  
-----