

-----  
Midmark M11 - v1.0.5  
Total Cycles: 691

Name: RITTER2  
Sterilizer ID: V2313228  
01/12/2022 04:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
2:00 189.3 F 4.3  
4:00 206.1 F 9.7  
6:00 222.1 F 6.9  
8:00 235.1 F 11.8  
10:00 236.3 F 10.2  
12:00 236.6 F 9.5  
14:00 251.7 F 16.8  
16:00 264.0 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.2  
1:00 272.4 F 29.7  
2:00 272.1 F 29.9  
3:00 271.6 F 29.1  
4:00 272.4 F 29.7  
  
Min 271.4 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.6 F 0.0  
5:00 225.0 F 0.0  
10:00 229.4 F 0.0  
15:00 230.3 F 0.0  
20:00 229.6 F 0.0  
25:00 229.9 F 0.0  
30:00 225.0 F 0.0  
Drying Complete

FILLING: 2:14  
HEATING: 17:45  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:06  
CYCLE COMPLETE

01/13/2022 07:30 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02