
Midmark M11 - v1.0.5
Total Cycles: 1578

Name: RITTER2
Sterilizer ID: V2313228
10/11/2022 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 98.9 F 0.0
4:00 106.5 F 0.0
6:00 137.0 F 0.0
8:00 174.1 F 0.0
10:00 198.1 F 0.2
12:00 212.3 F 3.8
14:00 222.9 F 7.3
16:00 232.8 F 11.1
18:00 231.7 F 7.2
20:00 242.1 F 12.4
22:00 239.4 F 10.3
24:00 236.7 F 8.8
26:00 251.0 F 15.8
28:00 262.7 F 22.6
30:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.0 F 28.8
2:00 272.4 F 29.2
3:00 272.7 F 29.3
4:00 272.7 F 29.3

Min 271.7 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 221.8 F 0.0
Dry Cycle Aborted

FILLING: 0:53
HEATING: 30:02
STERILIZING: 4:00
VENTING: 2:43
DRYING: 4:24
TOTAL CYCLE: 00:42:02

10/11/2022 09:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____