

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
2.1uS / 1.3ppm  
CYCLE NUMBER 001870  
11:08 18/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:58  
PRESSURIZING 2:04  
STERILIZING 3:11  
136.0 C 220kPa 3:11  
136.1 C 223kPa 3:41  
136.4 C 225kPa 4:11  
136.3 C 224kPa 4:41  
136.3 C 224kPa 5:11  
136.2 C 224kPa 5:41  
136.4 C 226kPa 6:11  
136.3 C 224kPa 6:41  
136.3 C 224kPa 6:41  
VENTING 6:41  
AIR DRYING 7:31  
CYCLE COMPLETE 7:42  
-----

Digital Signature #  
93089382106D2BA4  
-----