

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1154

Name: RITTER 1  
Sterilizer ID: V2178198  
09/09/2021 09:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 105.1 F 0.0  
2:00 107.8 F 0.0  
4:00 130.9 F 0.0  
6:00 161.3 F 0.9  
8:00 188.0 F 4.2  
10:00 202.8 F 8.5  
12:00 227.4 F 13.2  
14:00 228.7 F 7.7  
16:00 240.4 F 12.6  
18:00 237.5 F 9.8  
20:00 237.1 F 8.8  
22:00 248.7 F 14.6  
24:00 260.7 F 21.3  
26:00 271.4 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.4  
1:00 272.6 F 28.8  
2:00 272.3 F 28.9  
3:00 272.2 F 28.5  
4:00 272.6 F 29.1

Min 271.7 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.1 F 0.0  
5:00 213.7 F 0.0  
10:00 190.7 F 0.0  
15:00 187.5 F 0.0  
20:00 183.7 F 0.0  
25:00 180.9 F 0.0  
30:00 178.5 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 26:36  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:05:01  
CYCLE COMPLETE

09/09/2021 10:42 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02