
Midmark M11 - v1.0.5
Total Cycles: 2119

Name: RITTER2
Sterilizer ID: V2313228
03/06/2023 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 172.2 F 2.3
4:00 168.4 F 5.5
6:00 196.1 F 10.3
8:00 214.5 F 5.4
10:00 227.5 F 10.3
12:00 231.0 F 7.8
14:00 243.5 F 13.3
16:00 242.8 F 12.2
18:00 255.6 F 18.9
20:00 266.8 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 30.1
2:00 271.8 F 29.5
3:00 272.3 F 29.5
4:00 272.6 F 30.0

Min 271.0 F 28.8
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 230.7 F 0.0
10:00 220.9 F 0.0
15:00 218.1 F 0.0
20:00 219.3 F 0.0
25:00 220.4 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

03/06/2023 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____