

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 006975  
14:10 23/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:23  
PRESSURIZING 2:30  
STERILIZING 3:26  
136.0 C 220kPa 3:26  
136.1 C 222kPa 3:56  
136.4 C 225kPa 4:26  
136.1 C 221kPa 4:56  
136.1 C 223kPa 5:26  
136.1 C 220kPa 5:56  
136.4 C 223kPa 6:26  
136.4 C 224kPa 6:56  
136.4 C 224kPa 6:56  
VENTING 6:57  
AIR DRYING 7:45  
CYCLE COMPLETE 37:45  
-----

Digital Signature #  
2BB26D95E3C968A6  
-----