

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1537

Name: RITTER 1  
Sterilizer ID: V2178198  
01/19/2022 11:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.0 F 0.0  
2:00 196.9 F 3.2  
4:00 191.2 F 7.9  
6:00 218.7 F 13.3  
8:00 226.7 F 7.8  
10:00 239.4 F 13.1  
12:00 237.0 F 10.3  
14:00 237.6 F 8.1  
16:00 249.2 F 15.1  
18:00 261.0 F 21.7  
20:00 271.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.5  
1:00 272.6 F 29.2  
2:00 271.5 F 28.5  
3:00 272.6 F 29.1  
4:00 271.5 F 28.5  
  
Min 271.4 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 234.6 F 0.0  
10:00 230.1 F 0.0  
15:00 226.8 F 0.0  
20:00 227.0 F 0.0  
25:00 228.6 F 0.0  
30:00 230.1 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 20:25  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:04  
CYCLE COMPLETE

01/19/2022 12:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_