
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000903
15:14 03/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:03
STERILIZING 3:20
136.0 C 230kPa 3:20
136.4 C 230kPa 3:50
136.4 C 233kPa 4:20
136.3 C 231kPa 4:50
136.4 C 233kPa 5:20
136.4 C 233kPa 5:50
136.1 C 230kPa 6:20
136.5 C 231kPa 6:50
136.5 C 231kPa 6:50
VENTING 6:50
AIR DRYING 7:45
CYCLE COMPLETE 22:45

Digital Signature #
E1008BC1D631D3A3
