

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4048

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2023 02:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.5 F 0.0  
2:00 107.6 F 0.8  
4:00 124.6 F 2.9  
6:00 151.8 F 5.7  
8:00 180.4 F 9.3  
10:00 200.0 F 9.6  
12:00 217.9 F 5.1  
14:00 228.6 F 8.8  
16:00 238.4 F 13.0  
18:00 234.5 F 9.0  
20:00 242.7 F 11.6  
22:00 243.9 F 12.1  
24:00 255.5 F 18.1  
26:00 266.1 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.7 F 28.8  
2:00 271.7 F 28.4  
3:00 272.7 F 28.9  
4:00 271.6 F 28.2

Min 271.6 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 215.4 F 0.0  
10:00 228.5 F 0.0  
15:00 232.6 F 0.0  
20:00 226.3 F 0.0  
25:00 219.4 F 0.0  
30:00 212.7 F 0.0

Drying Complete

FILLING: 1:33  
HEATING: 27:42  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:06:49  
CYCLE COMPLETE

11/21/2023 03:29 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_